



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

**Rothe Enterprises, Inc. at
Metrology and Calibration Laboratory Services
(WMCLS) NASA / Wallops Flight Facility
34200 Fulton St., Building F-160
Wallops Island, VA 23337**

Fulfills the requirements of

ISO/IEC 17025:2017

and national standards

**ANSI/NCSL Z540-1-1994 (R2002) and
ANSI/NCSL Z540.3-2006 (R2013)**

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 21 May 2025

Certificate Number: AC-2986



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AND

ANSI/NCSL Z540-1-1994 (R2002)

ANSI/NCSL Z540.3-2006 (R2013)

**Rothe Enterprises, Inc. at
Metrology and Calibration Laboratory Services (WMCLS)**

NASA / Wallops Flight Facility

34200 Fulton St., Building F-160

Wallops Island, VA 23337

Edward Woodring 757-824-2368

CALIBRATION

Valid to: **May 21, 2025**

Certificate Number: **AC-2986**

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Source ¹	(0 to 0.22) V (0.22 to 2.2) V (2.2 to 11) V (11 to 22) V (22 to 220) V (220 to 1 100) V	9 μ V/V + 0.8 μ V 8 μ V/V + 1.2 μ V 8 μ V/V + 4 μ V 8 μ V/V + 8 μ V 9 μ V/V + 100 μ V 11 μ V/V + 600 μ V	Fluke 5700A Multifunction Calibrator with 5725A Amplifier
	Up to 330 mV 330 mV to 3.3 V (3.3 to 33) V (33 to 330) V 330 V to 1 kV	3 μ V/V + 1 μ V 2 μ V/V + 1.5 μ V 12 μ V/V + 20 μ V 18 μ V/V + 150 μ V 18 μ V/V + 1.5 mV	Fluke 5522A Multi-Product Calibrator
DC Voltage – Measure ¹	(0 to 100) mV 100 mV to 1 V (1 to 10) V (10 to 100) V (100 to 1 000) V	9 μ V/V + 300 nV 8 μ V/V + 300 nV 8 μ V/V + 500 nV 10 μ V/V + 30 μ V 22 μ V/V + 100 μ V	Hewlett Packard 3458A Digital Multimeter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Source ¹	Up to 220 μ A (0.22 to 2.2) mA (2.2 to 22) mA (22 to 220) mA (0.22 to 2.2) A (2.2 to 11) A	50 μ A/A + 8 nA 50 μ A/A + 8 nA 50 μ A/A + 80 nA 60 μ A/A + 0.8 μ A 80 μ A/A + 25 μ A 360 μ A/A + 480 μ A	Fluke 5700A Multifunction Calibrator with 5725A Amplifier
DC Current – Source ¹	(0 to 330) μ A 330 μ A to 3.3 mA (3.3 to 33) mA (33 to 330) mA 330 mA to 1.1 A (1.1 to 3.3) A (3.3 to 11) A (11 to 20.5) A	150 μ A/A + 0.02 μ A 100 μ A/A + 0.05 μ A 100 μ A/A + 0.25 μ A 100 μ A/A + 2.5 μ A 200 μ A/A + 40 μ A 380 μ A/A + 40 μ A 500 μ A/A + 500 μ A 1 mA/A + 750 μ A	Fluke 5522A Multi-Product Calibrator
DC Current – Measure ¹	(0 to 100) μ A 100 μ A to 1 mA (1 to 10) mA (10 to 100) mA 100 mA to 1 A	20 μ A/A + 800 pA 20 μ A/A + 5 nA 20 μ A/A + 50 nA 20 μ A/A + 500 nA 110 μ A/A + 10 μ A	Hewlett Packard 3458A Digital Multimeter
	(0 A to 100) A	39 μ A/A + 10 μ A	Hewlett Packard 3458A Digital Multimeter with Current Shunts
AC Voltage – Source ¹	220 μ V to 2.2 mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (2.2 to 22) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	550 μ V/V + 4.5 μ A 210 μ V/V + 4.5 μ V 105 μ V/V + 4.5 μ V 370 μ V/V + 4.5 μ V 850 μ V/V + 7 μ V 1.1 mV/V + 13 μ V 1.7 mV/V + 25 μ V 3.4 mV/V + 25 μ V 500 μ V/V + 5 μ V 210 μ V/V + 5 μ V 105 μ V/V + 5 μ V 370 μ V/V + 5 μ V 850 μ V/V + 7 μ V 1.1 mV/V + 12 μ V 1.7 mV/V + 25 μ V 3.4 mV/V + 25 μ V	Fluke 5700A Multifunction Calibrator with 5725A Amplifier



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source ¹	(22 to 220) mV		Fluke 5700A Multifunction Calibrator with 5725A Amplifier
	(10 to 20) Hz	550 $\mu\text{V}/\text{V}$ + 13 μV	
	(20 to 40) Hz	210 $\mu\text{V}/\text{V}$ + 8 μV	
	40 Hz to 20 kHz	105 $\mu\text{V}/\text{V}$ + 8 μV	
	(20 to 50) kHz	320 $\mu\text{V}/\text{V}$ + 8 μV	
	(50 to 100) kHz	850 $\mu\text{V}/\text{V}$ + 25 μV	
	(100 to 300) kHz	1.1 mV/V + 25 μV	
	(300 to 500) kHz	1.7 mV/V + 35 μV	
	500 kHz to 1 MHz	3.4 mV/V + 80 μV	
	220 mV to 2.2 V		
	(10 to 20) Hz	500 $\mu\text{V}/\text{V}$ + 80 μV	
	(20 to 40) Hz	160 $\mu\text{V}/\text{V}$ + 25 μV	
	40 Hz to 20 kHz	75 $\mu\text{V}/\text{V}$ + 6 μV	
	(20 to 50) kHz	120 $\mu\text{V}/\text{V}$ + 16 μV	
	(50 to 100) kHz	250 $\mu\text{V}/\text{V}$ + 70 μV	
	(100 to 300) kHz	500 $\mu\text{V}/\text{V}$ + 130 μV	
	(300 to 500) kHz	1.3 mV/V + 350 μV	
	500kHz to 1 MHz	2.7 mV/V + 850 μV	
	(2.2 to 22) V		
	(10 to 20) Hz	500 $\mu\text{V}/\text{V}$ + 800 μV	
	(20 to 40) Hz	160 $\mu\text{V}/\text{V}$ + 250 μV	
	40 Hz to 20 kHz	75 $\mu\text{V}/\text{V}$ + 60 μV	
	(20 to 50) kHz	120 $\mu\text{V}/\text{V}$ + 160 μV	
	(50 to 100) kHz	250 $\mu\text{V}/\text{V}$ + 350 μV	
	(100 to 300) kHz	500 $\mu\text{V}/\text{V}$ + 1.5 mV	
	(300 to 500) kHz	1.3 mV/V + 4.3 mV	
	500 kHz to 1 MHz	2.7 mV/V + 8.5 mV	
	(22 to 220) V		
(10 to 20) Hz	500 $\mu\text{V}/\text{V}$ + 8 mV		
(20 to 40) Hz	160 $\mu\text{V}/\text{V}$ + 2.5 mV		
40 Hz to 20 kHz	80 $\mu\text{V}/\text{V}$ + 0.8 mV		
(20 to 50) kHz	220 $\mu\text{V}/\text{V}$ + 3.5 mV		
(50 to 100) kHz	500 $\mu\text{V}/\text{V}$ + 8 mV		
(100 to 300) kHz	1.5 mV/V + 90 mV		
(300 to 500) kHz	4.7 mV/V + 90 mV		
500 kHz to 1 MHz	12 mV/V + 190 mV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source ¹	(220 to 1 100) V (15 to 50) Hz 50 Hz to 1 kHz 40 Hz to 1 kHz (1 to 20) kHz (20 to 30) kHz (220 to 750) V 30 Hz to 50 kHz (50 to 100) kHz	400 $\mu\text{V/V} + 16 \text{ mV}$ 80 $\mu\text{V/V} + 3.5 \text{ mV}$ 90 $\mu\text{V/V} + 4 \text{ mV}$ 165 $\mu\text{V/V} + 6 \text{ mV}$ 600 $\mu\text{V/V} + 11 \text{ mV}$ 600 $\mu\text{V/V} + 11 \text{ mV}$ 2.3 $\text{mV/V} + 45 \text{ mV}$	Fluke 5700A Multifunction Calibrator with 5725A Amplifier
AC Voltage – Source ¹	(1 to 33) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz (33 to 330) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz (0.33 to 3.3) V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz (3.3 to 33) V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz	800 $\mu\text{V/V} + 6 \mu\text{V}$ 150 $\mu\text{V/V} + 6 \mu\text{V}$ 200 $\mu\text{V/V} + 6 \mu\text{V}$ 1 $\text{mV/V} + 6 \mu\text{V}$ 3.5 $\text{mV/V} + 12 \mu\text{V}$ 8 $\text{mV/V} + 50 \mu\text{V}$ 300 $\mu\text{V/V} + 8 \mu\text{V}$ 145 $\mu\text{V/V} + 8 \mu\text{V}$ 160 $\mu\text{V/V} + 8 \mu\text{V}$ 350 $\mu\text{V/V} + 8 \mu\text{V}$ 800 $\mu\text{V/V} + 32 \mu\text{V}$ 2 $\text{mV/V} + 70 \mu\text{V}$ 300 $\mu\text{V/V} + 50 \mu\text{V}$ 150 $\mu\text{V/V} + 60 \mu\text{V}$ 190 $\mu\text{V/V} + 60 \mu\text{V}$ 300 $\mu\text{V/V} + 50 \mu\text{V}$ 700 $\mu\text{V/V} + 125 \mu\text{V}$ 2.4 $\text{mV/V} + 600 \mu\text{V}$ 300 $\mu\text{V/V} + 650 \mu\text{V}$ 150 $\mu\text{V/V} + 600 \mu\text{V}$ 240 $\mu\text{V/V} + 600 \mu\text{V}$ 350 $\mu\text{V/V} + 600 \mu\text{V}$ 900 $\mu\text{V/V} + 1.6 \text{ mV}$	Fluke 5522A Multi-Product Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source ¹	(33 to 330) V 45Hz to 1 kHz (1 to 10) kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz	190 μ V/V + 2 mV 200 μ V/V + 6 mV 250 μ V/V + 6 mV 300 μ V/V + 6 mV 2 mV/V + 50 mV	Fluke 5522A Multi-Product Calibrator
	(330 to 1 020) V 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	300 μ V/V + 10 mV 250 μ V/V + 10 mV 300 μ V/V + 10 mV	
AC Voltage – Measure ¹	Up to 10 mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz	300 μ V/V + 3 μ V 200 μ V/V + 1.1 μ V 300 μ V/V + 1.1 μ V 1 mV/V + 1.1 μ V 5 mV/V + 1.1 μ V 40 mV/V + 2 μ V	Hewlett Packard 3458A Digital Multimeter
	(10 to 100) mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz (0.1 to 1) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz (1 to 2) MHz	70 μ V/V + 4 μ V 70 μ V/V + 2 μ V 140 μ V/V + 2 μ V 300 μ V/V + 2 μ V 800 μ V/V + 2 μ V 3 mV/V + 10 μ V 10 mV/V + 10 μ V 15 mV/V + 10 μ V 70 μ V/V + 40 μ V 70 μ V/V + 20 μ V 140 μ V/V + 20 μ V 300 μ V/V + 20 μ V 800 μ V/V + 20 μ V 3 mV/V + 100 μ V 10 mV/V + 100 μ V 15 mV/V + 100 μ V	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure ¹	(1 to 10) V		Hewlett Packard 3458A Digital Multimeter
	(1 to 40) Hz	70 μ V/V + 400 μ V	
	40 Hz to 1 kHz	70 μ V/V + 200 μ V	
	(1 to 20) kHz	140 μ V/V + 200 μ V	
	(20 to 50) kHz	300 μ V/V + 200 μ V	
	(50 to 100) kHz	800 μ V/V + 200 μ V	
	(100 to 300) kHz	3 mV/V + 1 mV	
	300 kHz to 1 MHz	10 mV/V + 1 mV	
	(1 to 2) MHz	15 mV/V + 1 mV	
	(10 to 100) V		
	(1 to 40) Hz	200 μ V/V + 4 mV	
	40 Hz to 1 kHz	200 μ V/V + 2 mV	
	(1 to 20) kHz	200 μ V/V + 2 mV	
	(20 to 50) kHz	350 μ V/V + 2 mV	
	(50 to 100) kHz	1.2 mV/V + 2 mV	
	(100 to 300) kHz	4 mV/V + 10 mV	
	300 kHz to 1 MHz	15 mV/V + 10 mV	
	(100 to 1 000) V		
(1 to 40) Hz	400 μ V/V + 40 mV		
40 Hz to 1 kHz	400 μ V/V + 20 mV		
(1 to 20) kHz	600 μ V/V + 20 mV		
(20 to 50) kHz	1.2 mV/V + 20 mV		
(50 to 100) kHz	3 mV/V + 20 mV		
AC Voltage – Measure	Up to 2.2 mV		Fluke 5790A AC Measurement Standard
	(10 to 20) Hz	0.17 % of reading + 1.3 μ V	
	(20 to 40) Hz	0.074 % of reading + 1.3 μ V	
	40 Hz to 20 kHz	0.042 % of reading + 1.3 μ V	
	(20 to 50) kHz	0.082 % of reading + 2 μ V	
	(50 to 100) kHz	0.12 % of reading + 2.5 μ V	
	(100 to 300) kHz	0.23 % of reading + 4 μ V	
	(300 to 500) kHz	0.26 % of reading + 8 μ V	
	500 kHz to 1 MHz	0.5 % of reading + 8 μ V	
	(2.2 to 7) mV		
	(10 to 20) Hz	0.085 % of reading + 1.3 μ V	
	(20 to 40) Hz	0.037 % of reading + 1.3 μ V	
	40 Hz to 20 kHz	0.021 % of reading + 1.3 μ V	
	(20 to 50) kHz	0.041 % of reading + 2 μ V	
	(50 to 100) kHz	0.061 % of reading + 2.5 μ V	
	(100 to 300) kHz	0.12 % of reading + 4 μ V	
	(300 to 500) kHz	0.14 % of reading + 8 μ V	
	500 kHz to 1 MHz	0.36 % of reading + 8 μ V	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure	(7 to 22) mV		Fluke 5790A AC Measurement Standard
	(10 to 20) Hz	0.29 % of reading + 1.3 μ V	
	(20 to 40) Hz	0.019 % of reading + 1.3 μ V	
	40 Hz to 20 kHz	0.011 % of reading + 1.3 μ V	
	(20 to 50) kHz	0.021 % of reading + 2 μ V	
	(50 to 100) kHz	0.031 % of reading + 2.5 μ V	
	(100 to 300) kHz	0.082 % of reading + 4 μ V	
	(300 to 500) kHz	0.1 % of reading + 8 μ V	
	500 kHz to 1 MHz	0.26 % of reading + 8 μ V	
	(22 to 70) mV		
	(10 to 20) Hz	0.024 % of reading + 1.5 μ V	
	(20 to 40) Hz	0.013 % of reading + 1.5 μ V	
	40 Hz to 20 kHz	0.006 9 % of reading + 1.5 μ V	
	(20 to 50) kHz	0.013 % of reading + 2 μ V	
	(50 to 100) kHz	0.026 % of reading + 2.5 μ V	
	(100 to 300) kHz	0.053 % of reading + 4 μ V	
	(300 to 500) kHz	0.068 % of reading + 80 μ V	
	500 kHz to 1 MHz	0.013 % of reading + 8 μ V	
	(70 to 220) mV		
	(10 to 20) Hz	0.021 % of reading + 1.5 μ V	
	(20 to 40) Hz	0.008 7 % of reading + 1.5 μ V	
	40 Hz to 20 kHz	0.004 3 % of reading + 1.5 μ V	
	(20 to 50) kHz	0.007 3 % of reading + 2 μ V	
	(50 to 100) kHz	0.016 % of reading + 2.5 μ V	
	(100 to 300) kHz	0.028 % of reading + 4 μ V	
	(300 to 500) kHz	0.04 % of reading + 8 μ V	
	500 kHz to 1 MHz	0.12 % of reading + 8 μ V	
	(220 to 700) mV		
(10 to 20) Hz	0.021 % of reading + 1.5 μ V		
(20 to 40) Hz	0.007 8 % of reading + 1.5 μ V		
40 Hz to 20 kHz	0.003 8 % of reading + 1.5 μ V		
(20 to 50) kHz	0.005 6 % of reading + 2 μ V		
(50 to 100) kHz	0.008 4 % of reading + 2.5 μ V		
(100 to 300) kHz	0.021 % of reading + 4 μ V		
(300 to 500) kHz	0.034 % of reading + 8 μ V		
500 kHz to 1 MHz	0.12 % of reading + 8 μ V		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure	(0.7 to 2.2) V		Fluke 5790A AC Measurement Standard
	(10 to 20) Hz	0.02 % of reading	
	(20 to 40) Hz	0.0069 % of reading	
	40 Hz to 20 kHz	0.0029 % of reading	
	(20 to 50) kHz	0.0052 % of reading	
	(50 to 100) kHz	0.0076 % of reading	
	(100 to 300) kHz	0.02 % of reading	
	(300 to 500) kHz	0.031 % of reading	
	500 kHz to 1 MHz	0.12 % of reading	
	(2.2 to 7) V		
	(10 to 20) Hz	0.02 % of reading	
	(20 to 40) Hz	0.007 % of reading	
	40 Hz to 20 kHz	0.0029 % of reading	
	(20 to 50) kHz	0.0053 % of reading	
	(50 to 100) kHz	0.0088 % of reading	
	(100 to 300) kHz	0.022 % of reading	
	(300 to 500) kHz	0.047 % of reading	
	500 kHz to 1 MHz	0.15 % of reading	
	(7 to 22) V		
	(10 to 20) Hz	0.02 % of reading	
	(20 to 40) Hz	0.007 % of reading	
	40 Hz to 20 kHz	0.0031 % of reading	
	(20 to 50) kHz	0.0053 % of reading	
	(50 to 100) kHz	0.0085 % of reading	
	(100 to 300) kHz	0.022 % of reading	
	(300 to 500) kHz	0.047 % of reading	
	500 kHz to 1 MHz	0.015 % of reading	
(22 to 70) V			
(10 to 20) Hz	0.02 % of reading		
(20 to 40) Hz	0.0072 % of reading		
40 Hz to 20 kHz	0.0039 % of reading		
(20 to 50) kHz	0.0063 % of reading		
(50 to 100) kHz	0.011 % of reading		
(100 to 300) kHz	0.022 % of reading		
(300 to 500) kHz	0.051 % of reading		
500 kHz to 1 MHz	0.15 % of reading		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure	(70 to 220) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz (220 to 700) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (700 to 1 000) V (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz	0.02 % of reading 0.007 2 % of reading 0.003 8 % of reading 0.007 7 % of reading 0.011 % of reading 0.026 % of reading 0.07 % of reading 0.02 % of reading 0.011 % of reading 0.004 7 % of reading 0.015 % of reading 0.085 % of reading 0.02 % of reading 0.011 % of reading 0.004 4 % of reading 0.015 % of reading 0.085 % of reading	Fluke 5790A AC Measurement Standard
AC Voltage – Measure Flatness referenced to 1 kHz	Up to 2.2 mV (10 Hz to 30) Hz (30 Hz to 120) Hz 120 Hz to 1.2 kHz (1.2 to 120) kHz (120 to 500) kHz 500 kHz to 1.2 MHz (1.2 to 2) MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz (2.2 to 7) mV (10 Hz to 30) Hz (30 Hz to 120) Hz 120 Hz to 1.2 kHz (1.2 to 120) kHz (120 to 500) kHz 500 kHz to 1.2 MHz (1.2 to 2) MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	0.1 % of reading 0.05 % of reading 0.05 % of reading 0.05 % of reading 0.07 % of reading + 1 μ V 0.07 % of reading + 1 μ V 0.07 % of reading + 1 μ V 0.17 % of reading + 1 μ V 0.3 % of reading + 1 μ V 0.7 % of reading + 2 μ V 0.1 % of reading 0.05 % of reading 0.05 % of reading 0.05 % of reading 0.07 % of reading + 1 μ V 0.07 % of reading + 1 μ V 0.07 % of reading + 1 μ V 0.1 % of reading + 1 μ V 0.17 % of reading + 1 μ V 0.37 % of reading + 1 μ V	Fluke 5790A AC Measurement Standard



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure Flatness referenced to 1 kHz	(7 to 22) mV		Fluke 5790A AC Measurement Standard
	(10 Hz to 30) Hz	0.1 % of reading	
	(30 Hz to 120) Hz	0.05 % of reading	
	120 Hz to 1.2 kHz	0.05 % of reading	
	(1.2 to 120) kHz	0.05 % of reading	
	(120 to 500) kHz	0.07 % of reading	
	500 kHz to 1.2 MHz	0.07 % of reading	
	(1.2 to 2) MHz	0.07 % of reading	
	(2 to 10) MHz	0.1 % of reading	
	(10 to 20) MHz	0.17 % of reading	
	(20 to 30) MHz	0.37 % of reading	
	(22 to 70) mV		
	(10 Hz to 30) Hz	0.1 % of reading	
	(30 Hz to 120) Hz	0.05 % of reading	
	120 Hz to 1.2 kHz	0.05 % of reading	
	(1.2 to 120) kHz	0.05 % of reading	
	(120 to 500) kHz	0.05 % of reading	
	500 kHz to 1.2 MHz	0.05 % of reading	
	(1.2 to 2) MHz	0.05 % of reading	
	(2 to 10) MHz	0.1 % of reading	
	(10 to 20) MHz	0.15 % of reading	
	(20 to 30) MHz	0.35 % of reading	
	(70 to 220) mV		
	(10 Hz to 30) Hz	0.1 % of reading	
	(30 Hz to 120) Hz	0.04 % of reading	
	120 Hz to 1.2 kHz	0.04 % of reading	
	(1.2 to 120) kHz	0.04 % of reading	
(120 to 500) kHz	0.04 % of reading		
500 kHz to 1.2 MHz	0.05 % of reading		
(1.2 to 2) MHz	0.05 % of reading		
(2 to 10) MHz	0.1 % of reading		
(10 to 20) MHz	0.15 % of reading		
(20 to 30) MHz	0.35 % of reading		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure Flatness referenced to 1 kHz	(220 to 700) mV		Fluke 5790A AC Measurement Standard
	(10 Hz to 30) Hz	0.1 % of reading	
	(30 Hz to 120) Hz	0.03 % of reading	
	120 Hz to 1.2 kHz	0.03 % of reading	
	(1.2 to 120) kHz	0.03 % of reading	
	(120 to 500) kHz	0.03 % of reading	
	500 kHz to 1.2 MHz	0.05 % of reading	
	(1.2 to 2) MHz	0.05 % of reading	
	(2 to 10) MHz	0.1 % of reading	
	(10 to 20) MHz	0.15 % of reading	
	(20 to 30) MHz	0.35 % of reading	
	(0.7 to 2.2) V		
	(10 Hz to 30) Hz	0.1 % of reading	
	(30 Hz to 120) Hz	0.03 % of reading	
	120 Hz to 1.2 kHz	0.03 % of reading	
	(1.2 to 120) kHz	0.03 % of reading	
	(120 to 500) kHz	0.03 % of reading	
	500 kHz to 1.2 MHz	0.05 % of reading	
	(1.2 to 2) MHz	0.05 % of reading	
	(2 to 10) MHz	0.1 % of reading	
	(10 to 20) MHz	0.15 % of reading	
(20 to 30) MHz	0.35 % of reading		
(2.2 to 3.5) V			
(10 Hz to 30) Hz	0.1 % of reading		
(30 Hz to 120) Hz	0.03 % of reading		
120 Hz to 1.2 kHz	0.03 % of reading		
(1.2 to 120) kHz	0.03 % of reading		
(120 to 500) kHz	0.03 % of reading		
500 kHz to 1.2 MHz	0.05 % of reading		
(1.2 to 2) MHz	0.05 % of reading		
(2 to 10) MHz	0.1 % of reading		
(10 to 20) MHz	0.15 % of reading		
(20 to 30) MHz	0.35 % of reading		
AC Current – Source ¹	Up to 220 μ A		Fluke 5700A Multifunction Calibrator with 5725A Amplifier
	(10 to 20) Hz	700 μ A/A + 25 nA	
	(20 to 40) Hz	350 μ A/A + 20 nA	
	40 Hz to 1 kHz	140 μ A/A + 16 nA	
	(1 to 5) kHz	600 μ A/A + 40 nA	
	(5 to 10) kHz	1.6 mA/A + 80 nA	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source ¹	220 μ A to 2.2 mA (10 to 20) Hz (20 to 40) Hz 40Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	700 μ A/A + 40 nA 350 μ A/A + 35 nA 140 μ A/A + 35 nA 600 μ A/A + 400 nA 1.6 mA/A + 800 nA	Fluke 5700A Multifunction Calibrator with 5725A Amplifier
	(2.2 to 22) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (22 to 220) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz 220 mA to 2.2 A 20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (2.2 to 11) A 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	700 μ A/A + 400 nA 350 μ A/A + 350 nA 140 μ A/A + 350 nA 600 μ A/A + 4 μ A 1.6 mA/A + 8 μ A 700 μ A/A + 4 μ A 350 μ A/A + 3.5 μ A 140 μ A/A + 3.5 μ A 600 μ A/A + 40 μ A 1.6 mA/A + 80 μ A 650 μ A/A + 35 μ A 750 μ A/A + 80 μ A 8.5 mA/A + 160 μ A 460 μ A/A + 170 μ A 950 μ A/A + 380 μ A 3.6 mA/A + 750 μ A	
AC Current – Source ¹	(29 to 330) μ A (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz (0.33 to 3.3) mA (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	0.2 % of reading + 0.1 nA 0.15 % of reading + 0.1 nA 0.13 % of reading + 0.1 nA 0.3 % of reading + 0.15 nA 0.8 % of reading + 0.2 nA 1.6 % of reading + 0.4 nA 0.2 % of reading + 0.15 nA 0.13 % of reading + 0.15 nA 0.1 % of reading + 0.15 nA 0.2 % of reading + 0.2 nA 0.5 % of reading + 0.3 nA 1 % of reading + 0.6 nA	Fluke 5522A Multi-Product Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source ¹	(3.3 to 33) mA		Fluke 5522A Multi-Product Calibrator
	(10 to 20) Hz	0.18 % of reading + 2 nA	
	(20 to 45) Hz	0.09 % of reading + 2 nA	
	45 Hz to 1 kHz	0.04 % of reading + 2 nA	
	(1-5) kHz	0.08 % of reading + 2 nA	
	(5 to 10) kHz	0.2 % of reading + 3 nA	
	(10 to 30) kHz	0.4 % of reading + 4 nA	
	(33 to 330) mA		
	(10 to 20) Hz	0.18 % of reading + 20 nA	
	(20 to 45) Hz	0.09 % of reading + 20 nA	
	45 Hz to 1 kHz	0.04 % of reading + 20 nA	
	(1 to 5) kHz	0.1 % of reading + 50 nA	
	(5 to 10) kHz	0.2 % of reading + 100 nA	
	(10 to 30) kHz	0.4 % of reading + 200 nA	
	330 mA to 1.1 A		
	(10 to 45) Hz	0.18 % of reading + 100 nA	
45 Hz to 1 kHz	0.05 % of reading + 100 nA		
(1 to 5) kHz	0.6 % of reading + 1 μA		
(5 to 10) kHz	2.5 % of reading + 5 μA		
(3 to 11) A			
(45 to 100) Hz	0.06 % of reading + 2 μA		
100 Hz to 1 kHz	0.1 % of reading + 2 μA		
(1 to 5) kHz	3 % of reading + 2 μA		
(11 to 20.5) A			
(45 to 100) Hz	0.12 % of reading + 5 μA		
100 Hz to 1 kHz	0.15 % of reading + 5 μA		
(1 to 5) kHz	3 % of reading + 5 μA		
AC Current – Measure ¹	(5 to 100) μA		Hewlett Packard 3458A Digital Multimeter
	(10 to 20) Hz	0.4 % of reading + 30 nA	
	(20 to 45) Hz	0.15 % of reading + 30 nA	
	(45 to 100) Hz	0.06 % of reading + 30 nA	
	(100 Hz to 1 kHz)	0.06 % of reading + 30 nA	
	(0.1 to 1) mA		
	(10 to 20) Hz	0.4 % of reading + 200 nA	
	(20 to 45) Hz	0.15 % of reading + 200 nA	
	(45 to 100) Hz	0.06 % of reading + 200 nA	
	100 Hz to 5 kHz	0.03 % of reading + 200 nA	
	(5 to 20) kHz	0.06 % of reading + 200 nA	
	(20 to 50) kHz	0.4 % of reading + 400 nA	
(50 to 100) kHz	0.55 % of reading + 1.5 μA		

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure ¹	(1 to 10) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (5 to 20) kHz (20 to 50) kHz (50 to 100) kHz (10 to 100) mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100Hz to 5 kHz (5 to 20) kHz (20 to 50) kHz (50 to 100) kHz 100 mA to 1 A (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (5 to 20) kHz (20 to 50) kHz	0.4 % of reading + 2 μ A 0.15 % of reading + 2 μ A 0.06 % of reading + 2 μ A 0.03 % of reading + 2 μ A 0.06 % of reading + 2 μ A 0.4 % of reading + 4 μ A 0.55 % of reading + 15 μ A 0.4 % of reading + 20 μ A 0.15 % of reading + 20 μ A 0.06 % of reading + 20 μ A 0.03 % of reading + 20 μ A 0.06 % of reading + 20 μ A 0.4 % of reading + 40 μ A 0.55 % of reading + 150 μ A 0.4 % of reading + 200 μ A 0.16 % of reading + 200 μ A 0.08 % of reading + 200 μ A 0.1 % of reading + 200 μ A 0.3 % of reading + 200 μ A 1 % of reading + 400 μ A	Hewlett Packard 3458A Digital Multimeter
Resistance – Source ¹	0 Ω 1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 k Ω 1.9 k Ω 10 k Ω 19 k Ω 100 k Ω 190 k Ω 1 M Ω 1.9 M Ω 10 M Ω 19 M Ω 100 M Ω	40 $\mu\Omega$ 95 $\mu\Omega/\Omega$ 95 $\mu\Omega/\Omega$ 23 $\mu\Omega/\Omega$ 23 $\mu\Omega/\Omega$ 10 $\mu\Omega/\Omega$ 10 $\mu\Omega/\Omega$ 8.5 $\mu\Omega/\Omega$ 8.5 $\mu\Omega/\Omega$ 8.5 $\mu\Omega/\Omega$ 8.5 $\mu\Omega/\Omega$ 11 $\mu\Omega/\Omega$ 11 $\mu\Omega/\Omega$ 20 $\mu\Omega/\Omega$ 21 $\mu\Omega/\Omega$ 40 $\mu\Omega/\Omega$ 47 $\mu\Omega/\Omega$ 100 $\mu\Omega/\Omega$	Fluke 5522A Multi-Product Calibrator



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Source ¹	(0 to 11) Ω	40 μΩ/Ω + 0.01 Ω	Fluke 5522A Multi-Product Calibrator
	(11 to 33) Ω	30 μΩ/Ω + 0.015 Ω	
	(33 to 110) Ω	28 μΩ/Ω + 0.015 Ω	
	(110 to 330) Ω	28 μΩ/Ω + 0.02 Ω	
	(330 to 1 100) Ω	28 μΩ/Ω + 0.02 Ω	
	(1.1 to 3.3) kΩ	28 μΩ/Ω + 0.2 Ω	
	(3.3 to 11) kΩ	28 μΩ/Ω + 0.1 Ω	
	(11 to 33) kΩ	28 μΩ/Ω + 1 Ω	
	(33 to 110) kΩ	28 μΩ/Ω + 1 Ω	
	(110 to 330) kΩ	32 μΩ/Ω + 10 Ω	
	(330 to 1 100) kΩ	32 μΩ/Ω + 10 Ω	
	(1.1 to 3.3) MΩ	60 μΩ/Ω + 150 Ω	
	(3.3 to 11) MΩ	130 μΩ/Ω + 250 Ω	
	(11 to 33) MΩ	250 μΩ/Ω + 2500 Ω	
	(33 to 110) MΩ	500 μΩ/Ω + 3 kΩ	
(110 to 330) MΩ	3 mΩ/Ω + 100 kΩ		
(330 to 1 100) MΩ	15 mΩ/Ω + 500 kΩ		
Resistance – Source ¹	10 mΩ	0.4 μΩ	Fixed Resistors and HRRS Series 100G Ω High Resistance Decade Substituter
	1 Ω	0.2 μΩ	
	10 kΩ	4.1 mΩ	
	1 MΩ	10 Ω	
	2 MΩ	41 Ω	
	3 MΩ	61 Ω	
	4 MΩ	81 Ω	
	5 MΩ	100 Ω	
	6 MΩ	120 Ω	
	7 MΩ	140 Ω	
	8 MΩ	160 Ω	
	9 MΩ	180 Ω	
	10 MΩ	200 Ω	
	12 MΩ	2 Ω	
	13 MΩ	3 Ω	
	14 MΩ	4.1 Ω	
	15 MΩ	5.1 Ω	
	16 MΩ	6.1 kΩ	
	17 MΩ	7.1 kΩ	
	18 MΩ	8.1 kΩ	
19 MΩ	9.1 kΩ		
100 MΩ	10 kΩ		
200 MΩ	71 kΩ		
300 MΩ	110 kΩ		
400 MΩ	140 kΩ		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Source ¹	500 MΩ 600 MΩ 700 MΩ 800 MΩ 900 MΩ 1.0 GΩ 2.0 GΩ 3.0 GΩ 4.0 GΩ 5.0 GΩ 6.0 GΩ 7.0 GΩ 8.0 GΩ 9.0 GΩ 10.0 GΩ	180 kΩ 210 kΩ 250 kΩ/ Ω 280 kΩ/ Ω 320 kΩ/ Ω 510 kΩ/ Ω 1 MΩ/ Ω 1.5 MΩ/ Ω 2 MΩ/ Ω 2.5 MΩ/ Ω 3 MΩ/ Ω 3.6 MΩ/ Ω 4.1 MΩ/ Ω 4.6 MΩ/ Ω 7.1 MΩ/ Ω	Fixed Resistors and HRRS Series 100G Ω High Resistance Decade Substituter
Resistance – Measure ¹	0 to 10 Ω Up to 100 Ω Up to 1 kΩ Up to 10 kΩ Up to 100 kΩ Up to 1 MΩ Up to 10 MΩ Up to 100 MΩ Up to 1 GΩ	15 μΩ/Ω + 50 μΩ 12 μΩ/Ω + 500 μΩ 10 μΩ/Ω + 500 μΩ 10 μΩ/Ω + 5 mΩ 10 μΩ/Ω + 50 mΩ 15 μΩ/Ω + 2 Ω 50 μΩ/Ω + 100 Ω 500 μΩ/Ω + 1 kΩ 0.5% of reading + 10 kΩ	Hewlett Packard 3458A Digital Multimeter
Capacitance – Source ¹	(220 to 400) pF 0.4 pF to 1.1 nF (1.1 to 3.3) nF (3.3 to 11) nF (11 to 33) nF (33 to 110) nF (110 to 330) nF 330 nF to 1.1 μF (1.1 to 3.3) μF (3.3 to 11) μF	0.5 % of reading + 10 pF 0.5 % of reading + 0.01 nF 0.5 % of reading + 0.01 nF 0.25 % of reading + 0.01 nF 0.25 % of reading + 0.1 nF 0.25 % of reading + 0.1 nF 0.25 % of reading + 0.3 nF 0.25 % of reading + 1 nF 0.25 % of reading + 3 nF 0.25 % of reading + 10 nF	Fluke 5522A Multi-Product Calibrator
Capacitance – Source ¹	(11 to 33) μF (33 to 110) μF (110 to 330) μF 330 μF to 1.1 mF (1.1 to 3.3) mF (3.3 to 11) mF (11 to 33) mF (33 to 110) mF	0.4 % of reading + 30 nF 0.45 % of reading + 100 nF 0.45 % of reading + 300 nF 0.45 % of reading + 1 μF 0.45 % of reading + 3 μF 0.45 % of reading + 10 μF 0.75 % of reading + 30 μF 1.1 % of reading + 100 μF	Fluke 5522A Multi-Product Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Inductance – Source ¹	100 μ H		Fixed Inductors
	100 Hz	240 μ H	
	200 Hz	240 μ H	
	400 Hz	240 μ H	
	1 kHz	240 μ H	
	10 kHz	240 μ H	
	1 mH		
	100 Hz	240 μ H	
	200 Hz	240 μ H	
	400 Hz	240 μ H	
	1 kHz	240 μ H	
	10 kHz	240 μ H	
	1 H		
	100 Hz	270 μ H	
	200 Hz	270 μ H	
	400 Hz	270 μ H	
1 kHz	270 μ H		
10 H			
100 Hz	1.2mH		
200 Hz	1.2mH		
400 Hz	1.2mH		
1 kHz	1.2mH		
Electrical Simulation of Thermocouple Indicators Source and Measure ¹	Type B		Fluke 5522A Multi-Product Calibrator
	(600 to 800) °C	0.44 °C	
	(800 to 1 000) °C	0.34 °C	
	(1 000 to 1 550) °C	0.3 °C	
	(1 550 to 1 820) °C	0.33 °C	
	Type C		
	(0 to 150) °C	0.3 °C	
	(150 to 650) °C	0.26 °C	
	(650 to 1 000) °C	0.31 °C	
	(1 000 to 1 800) °C	0.5 °C	
	(1 800 to 2 316) °C	0.84 °C	
	Type E		
	(-250 to -100) °C	0.5 °C	
	(-100 to -25) °C	0.16 °C	
(-25 to +350) °C	0.14 °C		
(350 to 650) °C	0.16 °C		
(650 to 1 000) °C	0.21 °C		

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicators Source and Measure ¹	Type J		Fluke 5522A Multi-Product Calibrator
	(-210 to -100) °C	0.27 °C	
	(-100 to -30) °C	0.16 °C	
	(-30 to +150) °C	0.14 °C	
	(150 to 760) °C	0.17 °C	
	(760 to 1 200) °C	0.23 °C	
	Type K		
	(-200 to -100) °C	0.33 °C	
	(-100 to -25) °C	0.18 °C	
	(-25 to +120) °C	0.16 °C	
	(120 to 1 000) °C	0.26 °C	
	(1 000 to 1 372) °C	0.4 °C	
	Type L		
	(-200 to -100) °C	0.37 °C	
	(-100 to +800) °C	0.26 °C	
	(800 to 900) °C	0.17 °C	
	Type N		
	(-200 to -100) °C	0.4 °C	
	(-100 to -25) °C	0.22 °C	
	(-25 to +120) °C	0.19 °C	
	(120 to 410) °C	0.18 °C	
	(410 to 1 300) °C	0.27 °C	
	Type R		
	(0 to 250) °C	0.57 °C	
	(250 to 400) °C	0.35 °C	
	(400 to 1 000) °C	0.33 °C	
	(1 000 to 1 767) °C	0.4 °C	
	Type S		
(0 to 250) °C	0.47 °C		
(250 to 1 000) °C	0.36 °C		
(1 000 to 1 400) °C	0.37 °C		
(1 400 to 1 767) °C	0.46 °C		
Type T			
(-250 to -150) °C	0.63 °C		
(-150 to 0) °C	0.24 °C		
(0 to 120) °C	0.16 °C		
(120 to 400) °C	0.14 °C		
Type U			
(-200 to 0) °C	0.56 °C		
(0 to 600) °C	0.27 °C		

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of RTD Indicating Devices ¹	PT385-100 Ω		Fluke 5522A Multi-Product Calibrator
	(-200 to -80) °C	0.05 °C	
	(-80 to 0) °C	0.05 °C	
	(0 to 100) °C	0.07 °C	
	(100 to 300) °C	0.09 °C	
	(300 to 400) °C	0.1 °C	
	(400 to 630) °C	0.12 °C	
	(630 to 800) °C	0.23 °C	
	PT3926-100 Ω		
	(-200 to -80) °C	0.05 °C	
	(-80 to 0) °C	0.05 °C	
	(0 to 100) °C	0.07 °C	
	(100 to 300) °C	0.09 °C	
	(300 to 400) °C	0.1 °C	
	(400 to 630) °C	0.12 °C	
	PT3916-100 Ω		
	(-200 to -190) °C	0.25 °C	
	(-190 to -80) °C	0.04 °C	
	(-80 to 0) °C	0.05 °C	
	(0 to 100) °C	0.06 °C	
	(100 to 260) °C	0.07 °C	
	(260 to 300) °C	0.08 °C	
	(300 to 400) °C	0.09 °C	
	(400 to 600) °C	0.1 °C	
	(600 to 630) °C	0.23 °C	
	PT385-200 Ω		
	(-200 to -80) °C	0.04 °C	
	(-80 to 0) °C	0.04 °C	
(0 to 100) °C	0.04 °C		
(100 to 260) °C	0.05 °C		
(260 to 300) °C	0.12 °C		
(300 to 400) °C	0.13 °C		
(400 to 600) °C	0.14 °C		
(600 to 630) °C	0.16 °C		

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of RTD Indicating Devices ¹	PT385-500 Ω		Fluke 5522A Multi-Product Calibrator
	(-200 to -80) °C	0.04 °C	
	(-80 to 0) °C	0.05 °C	
	(0 to 100) °C	0.05 °C	
	(100 to 260) °C	0.06 °C	
	(260 to 300) °C	0.08 °C	
	(300 to 400) °C	0.08 °C	
	(400 to 600) °C	0.09 °C	
	(600 to 630) °C	0.11 °C	
	PT385-1 000 Ω		
	(-200 to -80) °C	0.03 °C	
	(-80 to 0) °C	0.03 °C	
	(0 to 100) °C	0.04 °C	
	(100 to 260) °C	0.05 °C	
	(260 to 300) °C	0.06 °C	
(300 to 400) °C	0.07 °C		
(400 to 600) °C	0.07 °C		
(600 to 630) °C	0.23 °C		
PtNi 385-120 Ω			
(-80 to 0) °C	0.08 °C		
(0 to 100) °C	0.08 °C		
(100 to 260) °C	0.14 °C		
Electrical Simulation of RTD Indicating Devices ¹	Cu 427-10 Ω (-100 to +260) °C	0.3 °C	Fluke 5522A Multi-Product Calibrator
Oscilloscopes ¹ Amplitude	1 mV to 130 VDC (into 1 MΩ) (into 50 Ω)	0.05 % of reading + 40 μV 0.25 % of reading + 40 μV	Fluke 5522A SC1100 Multi-Product Calibrator
Oscilloscopes ¹ Square Wave, 10 Hz to 10 kHz	1 mV to 130 V, 1 MΩ 1 mv to 6.6 V, 50 Ω	0.1 % of reading + 40 μV 0.25 % of reading + 40 μV	
Oscilloscopes ¹ Waveform Generator (Square, Sine, Triangle)	1.8 mV to 55 Vpp, 1 MΩ 1.8 mV to 2.5 Vpp, 50 Ω	3 % of reading + 100 μV	
Oscilloscopes ¹ Pulse Generator	200 ns to 20 ms	5 % of Pulse Width reading + 2 ns	
Oscilloscopes ¹ Frequency Accuracy	0.2 Hz to 1 GHz	2.5 x 10 ⁻⁶ of reading	



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes ¹ Leveled Sinewave Absolute	50 kHz Reference 50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz (600 to 1 100) MHz	2 % of reading + 300 μV 3.5 % of reading + 300 μV 4 % of reading + 300 μV 6 % of reading + 300 μV 7 % of reading + 300 μV	Fluke 5522A SC1100 Multi-Product Calibrator
Oscilloscopes ¹ Leveled Sinewave Flatness (relative to 50 kHz)	50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz (600 to 1 100) MHz	1.5 % of reading + 100 μV 2 % of reading + 100 μV 4 % of reading + 100 μV 5 % of reading + 100 μV	
Oscilloscopes ¹ Impedance Measurement	40 to 60 Ω 500 kΩ to 1.5 MΩ	0.1 % of reading	
Oscilloscopes ¹ Capacitance Measurement	5 pF to 50 pF	5 % of reading + 0.5 pF	

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Source (Power Sensor or Thermistor Mount Calibration Factors)	10 μW to 10 mW 0.006 MHz 0.009 MHz (0.02 to 0.09) MHz 0.1 MHz 0.2 MHz (0.3 to 3) MHz 5 MHz (10 to 80) MHz (90 to 150) MHz (200 to 600) MHz 650 MHz 700 MHz to 2 GHz (2.1 to 3.6) GHz (3.7 to 4.6) GHz (4.8 to 10) GHz 10.2 GHz (10.4 to 10.6) GHz (10.8 to 11) GHz (11.2 to 11.8) GHz 12 GHz	0.52 % of reading 0.47 % of reading 0.44 % of reading 0.75 % of reading 0.71 % of reading 0.69 % of reading 0.7 % of reading 0.68 % of reading 0.69 % of reading 0.68 % of reading 0.69 % of reading 0.68 % of reading 0.7 % of reading 0.72 % of reading 0.77 % of reading 0.82 % of reading 0.83 % of reading 0.8 % of reading 0.81 % of reading 0.88 % of reading	Tegam 1830A or Agilent E4418B Power Meter with 2505A RF Power Standard



ANSI National Accreditation Board

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Power – Source (Power Sensor or Thermistor Mount Calibration Factors)	10 μ W to 10 mW (12.2 to 12.75) GHz (13 to 13.75) GHz (14 to 15) GHz 15.25 GHz (15.5 to 16.5) GHz (16.75 to 18) GHz	0.89 % of reading 0.87 % of reading 0.88 % of reading 0.89 % of reading 0.87 % of reading 0.88 % of reading	Tegam 1830A or Agilent E4418B Power Meter with 2505A RF Power Standard
RF Power – Source (Power Sensor or Thermistor Mount Calibration Factors)	10 μ W to 10 mW 10 MHz 20 MHz 30 MHz (40 to 90) MHz (100 to 250) MHz (300 to 350) MHz (400 to 450) MHz 500 MHz (550 to 650) MHz 700 MHz 750 MHz to 2.4 GHz 2.5 GHz (2.6 to 3) GHz (3.1 to 5.4) GHz (5.6 to 6.4) GHz (6.6 to 9.2) GHz (9.4 to 13.5) GHz (13.75 to 17.5) GHz 17.75 GHz 18 GHz 19 GHz 20 GHz 21 GHz 22 GHz 23 GHz (24 to 26.5) GHz	1.8 % of reading 1.3 % of reading 1.2 % of reading 1.2 % of reading 1.1 % of reading 1.1 % of reading 1.1 % of reading 1.1 % of reading 1.1 % of reading 1.2 % of reading 1.2 % of reading 1.2 % of reading 1.2 % of reading 1.3 % of reading 1.4 % of reading 1.5 % of reading 1.8 % of reading 2.3 % of reading 2.4 % of reading 2.4 % of reading 2.8 % of reading 2.5 % of reading 2.3 % of reading 2.5 % of reading 3 % of reading 2.9 % of reading	Tegam 1830A or Agilent E4418B Power Meter with F1135B RF Power Standard



ANSI National Accreditation Board

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Absolute Power – Source ^{1,2}	(>+20 to +24) dBm		Absolute Power -- Fluke 96040A Low Phase Noise Reference Source with 50 ohm Leveling Head
	10 Hz to 99.9 kHz	0.069 %X	
	100 kHz	0.069 %X	
	100 kHz to 10 MHz	0.071 %X	
	(10 to 128) MHz	0.071 %X	
	(128 to 300) MHz	0.071 %X	
	300 MHz to 1.4 GHz	0.087 %X	
	(1.4 to 3) GHz	0.28 %X	
	(3 to 4) GHz	0.96 %X	
	(>+14 to +20) dBm		
	10 Hz to 99.9 kHz	0.069 %X	
	100 kHz	0.069 %X	
	100 kHz to 10 MHz	0.071 %X	
	(10 to 128) MHz	0.071 %X	
	(128 to 300) MHz	0.071 %X	
	300 MHz to 1.4 GHz	0.087 %X	
	(1.4 to 3) GHz	0.28 %X	
	(3 to 4) GHz	0.96 %X	
	(-17 to + 14) dBm		
	10 Hz to 99.9 kHz	Greater of 0.003 dB or 0.069 %X	
	100 kHz	Greater of 0.003 dB or 0.069 %X	
	100 kHz to 10 MHz	Greater of 0.003 1 dB or 0.071 %X	
	(10 to 128) MHz	Greater of 0.003 1 dB or 0.071 %X	
	(128 to 300) MHz	Greater of 0.003 1 dB or 0.071 %X	
300 MHz to 1.4 GHz	Greater of 0.003 8 dB or 0.087 %X		
(1.4 to 3) GHz	Greater of 0.012 dB or 0.28 %X		
(3 to 4) GHz	Greater of 0.042 dB or 0.96 %X		
(-48 to <-17) dBm			
10 Hz to 99.9 kHz	0.069 %X		
100 kHz	0.069 %X		
100 kHz to 10 MHz	0.071 %X		
(10 to 128) MHz	0.071 %X		
(128 to 300) MHz	0.071 %X		
300 MHz to 1.4 GHz	0.087 %X		
(1.4 to 3) GHz	0.28 %X		
(3 to 4) GHz	0.99 %X		

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Absolute Power – Source ¹	(>-74 to <-48) dBm		Absolute Power -- Fluke 96040A Low Phase Noise Reference Source with 50 ohm Leveling Head
	10 Hz to 99.9 kHz	0.069 %X	
	100 kHz	0.087 %X	
	100 kHz to 10 MHz	0.087 %X	
	(10 to 128) MHz	0.074 %X	
	(128 to 300) MHz	0.074 %X	
	300 MHz to 1.4 GHz	0.13 %X	
	(1.4 to 3) GHz	0.3 %X	
	(3 to 4) GHz	0.99 %X	
	(>-84 to -74) dBm		
	10 Hz to 99.9 kHz	0.069 %X	
	100 kHz	0.15 %X	
	100 kHz to 10 MHz	0.15 %X	
	(10 to 128) MHz	0.074 %X	
	(128 to 300) MHz	0.11 %X	
	300 MHz to 1.4 GHz	0.15 %X	
	(1.4 to 3) GHz	0.37 %X	
	(3 to 4) GHz	1 %X	
	(>-94 to -84) dBm		
	10 Hz to 99.9 kHz	0.069 %X	
	100 kHz	0.15 %X	
	100 kHz to 10 MHz	0.15 %X	
	(10 to 128) MHz	0.11 %X	
	(128 to 300) MHz	0.15 %X	
300 MHz to 1.4 GHz	0.28 %X		
(1.4 to 3) GHz	0.37 %X		
(3 to 4) GHz	0.96 %X		
(-130 to -94) dBm			
10 Hz to 99.9 kHz	0.069 %X		
100 kHz	0.069 %X		
100 kHz to 10 MHz	0.069 %X		
(10 to 128) MHz	0.2 %X		
(128 to 300) MHz	0.41 %X		
300 MHz to 1.4 GHz	0.41 %X		
(1.4 to 3) GHz	0.48 %X		
(3 to 4) GHz	0.96 %X		



ANSI National Accreditation Board

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Power Flatness – Source ¹	(>+20 to +24) dBm		Fluke 96040A Low Phase Noise Reference Source with 50 Ω Leveling Head Relative to 100 kHz
	10 Hz to 99.9 kHz	0.003 dB	
	100 kHz to 10 MHz	0.003 1 dB	
	(10 to 128) MHz	0.003 1 dB	
	(128 to 300) MHz	0.003 dB	
	300 MHz to 1.4 GHz	0.003 dB	
	(1.4 to 3) GHz	0.012 dB	
	(3 to 4) GHz	0.042 dB	
	(>+14 to +20) dBm		
	10 Hz to 99.9 kHz	0.003 dB	
	100 kHz to 10 MHz	0.003 1 dB	
	(10 to 128) MHz	0.003 1 dB	
	(128 to 300) MHz	0.003 1 dB	
	300 MHz to 1.4 GHz	0.003 8 dB	
	(1.4 to 3) GHz	0.012 dB	
	(3 to 4) GHz	0.042 dB	
	(-17 to + 14) dBm		
	10 Hz to 99.9 kHz	0.003 dB	
	100 kHz to 10 MHz	0.003 1 dB	
	(10 to 128) MHz	0.003 1 dB	
	(128 to 300) MHz	0.003 1 dB	
	300 MHz to 1.4 GHz	0.003 8 dB	
	(1.4 to 3) GHz	0.012 dB	
	(3 to 4) GHz	0.042 dB	
	(-48 to <-17) dBm		
	10 Hz to 99.9 kHz	0.003 dB	
	100 kHz to 10 MHz	0.003 8 dB	
	(10 to 128) MHz	0.003 2 dB	
(128 to 300) MHz	0.003 2 dB		
300 MHz to 1.4 GHz	0.005 6 dB		
(1.4 to 3) GHz	0.013 dB		
(3 to 4) GHz	0.043 dB		
(>-74 to <-48) dBm			
10 Hz to 99.9 kHz	0.003 dB		
100 kHz to 10 MHz	0.003 8 dB		
(10 to 128) MHz	0.003 2 dB		
(128 to 300) MHz	0.003 2 dB		
300 MHz to 1.4 GHz	0.005 6 dB		
(1.4 to 3) GHz	0.013 dB		
(3 to 4) GHz	0.043 dB		



ANSI National Accreditation Board

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Power Flatness – Source ¹	(>-84 to -74) dBm		Fluke 96040A Low Phase Noise Reference Source with 50 Ω Leveling Head Relative to 100kHz
	10 Hz to 99.9 kHz	0.003 dB	
	100 kHz to 10 MHz	0.006 6 dB	
	(10 to 128) MHz	0.003 2 dB	
	(128 to 300) MHz	0.004 6 dB	
	300 MHz to 1.4 GHz	0.006 6 dB	
	(1.4 to 3) GHz	0.016 dB	
	(3 to 4) GHz	0.044 dB	
	(>-94 to -84) dBm		
	10 Hz to 99.9 kHz	0.003 dB	
	100 kHz to 10 MHz	0.006 6 dB	
	(10 to 128) MHz	0.004 6 dB	
	(128 to 300) MHz	0.006 6 dB	
	300 MHz to 1.4 GHz	0.012 dB	
	(1.4 to 3) GHz	0.016 dB	
(3 to 4) GHz	0.042 dB		
Attenuation – Source ¹ Relative to +16 dBm output	(-130 to -94) dBm		Fluke 96040A Low Phase Noise Reference Source with 50 Ω Leveling Head Relative to 100 kHz
	10 Hz to 99.9 kHz	0.003 dB	
	100 kHz to 10 MHz	0.003 dB	
	(10 to 128) MHz	0.003 dB	
	(128 to 300) MHz	0.003 dB	
	300 MHz to 1.4 GHz	0.003 dB	
(1.4 to 3) GHz	0.012 dB		
(3 to 4) GHz	0.42 dB		
Attenuation – Source ¹ Relative to +10 dBm output	(0 to 55) dB	0.29 dB	Fluke 96040A Low Phase Noise Reference Source with 50 Ω Leveling Head Relative to 100 kHz
	(55 to 64) dB	0.035 dB	
	(64 to 74) dB	0.059 dB	
	(74 to 100) dB	0.082 dB	
	(100 to 116) dB	0.18 dB	
Attenuation – Source ¹ Relative to +10 dBm output	(0 to 49) dB	0.024 dB	Fluke 96040A Low Phase Noise Reference Source with 50 Ω Leveling Head Relative to 100 kHz
	(49 to 59) dB	0.035 dB	
	(59 to 69) dB	0.059 dB	
	(69 to 94) dB	0.082 dB	
	(94 to 110) dB	0.18 dB	



ANSI National Accreditation Board

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AM Depth – Source ¹	99% AM Depth 10 % Depth 20 % Depth 40 % Depth 50 % Depth 60 % Depth 80 % Depth 90 % Depth	3.7 % Depth 3.7 % Depth 3.8 % Depth 3.8 % Depth 3.7 % Depth 3.8 % Depth 3.7 % Depth	Fluke 96040A Low Phase Noise Reference Source
FM Modulation – Source ¹	50 kHz range, 12.5 MHz Carrier, 100 kHz Deviation Modulation Rate: 20 Hz 400 Hz 1 kHz 10 kHz 50 kHz 75 kHz 100 kHz 125 kHz 150 kHz 200 kHz 50 kHz range, 400 MHz Carrier, 10 kHz Deviation Modulation Rate: 20 Hz 400 Hz 1 kHz 10 kHz 50 kHz 75 kHz 100 kHz 125 kHz 150 kHz 200 kHz	3.7 % Deviation 3.7 % Deviation 3.7 % Deviation 3.7 % Deviation 4.2 % Deviation 4.2 % Deviation 5.1 % Deviation 5.1 % Deviation 5.1 % Deviation 5.1 % Deviation 3.7 % Deviation 3.7 % Deviation 3.7 % Deviation 3.7 % Deviation 4.2 % Deviation 4.2 % Deviation 5.1 % Deviation 5.1 % Deviation 5.1 % Deviation 5.1 % Deviation	Fluke 96040A Low Phase Noise Reference Source



ANSI National Accreditation Board

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
FM Modulation – Source ¹	50 kHz range, 400 MHz Carrier, 500 kHz Deviation Modulation Rate: 20 Hz 400 Hz 1 kHz 10 kHz 50 kHz 75 kHz 100 kHz 125 kHz 150 kHz 200 kHz	3.8 % Deviation 3.8 % Deviation 3.8 % Deviation 3.8 % Deviation 4.2 % Deviation 4.2 % Deviation 5.2 % Deviation 5.2 % Deviation 5.2 % Deviation 5.2 % Deviation	FM -- Fluke 96040A Low Phase Noise Reference Source
RF Power – Measure Power Meter Reference at Fixed Frequencies	1 mW, 0 dBm (50, 100, 500) MHz 1 GHz	0.88 % of reading	TEGAM 1830A Power Meter with HP 478A-H75 Thermistor Mount

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pressure Source	(0.2 to 25) psi (20 to 1 000) psi	0.005 % of reading + 0.000 85 psi 0.005 % of reading + 0.026 psi	Ruska 2465, Ruska 2465-799 Mass Set, N ₂ medium with Pistons 2465-725 2465-729
	(500 to 15 000) psi	0.007 5 % of reading + 0.15 psi	Ruska 2475 with 2475-701 Piston and 2475-940 Mass Set, Helium Medium

Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency - Source	10 MHz	1×10^{-11} Hz/Hz	Xli GPS Receiver

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95% of reading.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. X = selected output value.
3. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-2986.



Jason Stine, Vice President

